



# WEBER SCIENTIFIC

## CERTIFICATE OF ANALYSIS

PRODUCT DESCRIPTION: Buffered Phosphate Dilution Water Item No.: 3024-25  
 VOLUME 5 gallon LOT No. 171701 EXP. DATE 06/30/17

TEST	RESULTS	SPECIFICATION	METHOD
pH	7.3	7.2 ± 0.2	WS cSOP 8.0
TOXICITY *	Count at 0 minutes: 75 Count at 45 minutes: 82	< 20 % decrease	WS cSOP 24.3
FILL VOLUME	CONFORMS	LABEL	WS SOP CP3024-25.2.2
COLOR	CONFORMS	CLEAR, COLORLESS, NO FOREIGN MATTER	VISUAL
PACKAGE INTEGRITY JERRYCAN	CONFORMS	NO LOOSE CAPS NO LEAKAGE	VISUAL
PACKAGE MARKINGS	CONFORMS	CONFORMS TO WS STD	WS SOP CP3024-25.4.0
WATER QUALITY	CONFORMS	MICROBIO- LOGICALLY SUITABLE	<i>SMEWW</i> TABLE 9020:II

Reference: *Standard Methods for the Examination of Dairy Products*, 17<sup>th</sup> edition, 2004, 4.020 – 4.042; *Standard Methods for the Examination of Water and Wastewater*, 20<sup>th</sup> edition

\*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories

Made By:  
Weber Scientific  
2732 Kuser Road  
Hamilton, NJ 08691  
(609)-584-7677

Fred Weber  
President

Date July 13, 2017