



WEBER SCIENTIFIC

CERTIFICATE OF ANALYSIS

PRODUCT DESCRIPTION: Buffered Phosphate Dilution Water Item No.: 3024-25
 VOLUME 5 gallon LOT No. 171504 EXP. DATE 12/2016

TEST	RESULTS	SPECIFICATION	METHOD
pH	7.3	7.2 ± 0.2	WS cSOP 8.0
TOXICITY *	Count at 0 minutes: 48 Count at 45 minutes: 50	< 20 % decrease	WS cSOP 24.3
FILL VOLUME	CONFORMS	LABEL	WS SOP CP3024-25.2.2
COLOR	CONFORMS	CLEAR, COLORLESS, NO FOREIGN MATTER	VISUAL
PACKAGE INTEGRITY JERRYCAN	CONFORMS	NO LOOSE CAPS NO LEAKAGE	VISUAL
PACKAGE MARKINGS	CONFORMS	CONFORMS TO WS STD	WS SOP CP3024-25.4.0
WATER QUALITY	CONFORMS	MICROBIO- LOGICALLY SUITABLE	<i>SMEWW</i> TABLE 9020:II
Reference: <i>Standard Methods for the Examination of Dairy Products</i> , 17 th edition, 2004, 4.020 – 4.042; <i>Standard Methods for the Examination of Water and Wastewater</i> , 20 th edition			

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories

Made By:
 Weber Scientific
 2732 Kuser Road
 Hamilton, NJ 08691
 (609)-584-7677

Fred Weber
 President

Date December 18, 2015