



2732 Kuser Road, Hamilton, NJ 08691
 800-328-8378
 www.weberscientific.com

How Often Should the Air Sampler be Recalibrated?

The Microbiologique Air Sampler ships calibrated and ready to use. For best results, we recommend recalibration once every **12 months**.

What is included with Recalibration?

Every recalibration includes a calibration certification for your records.

Additional Questions? Contact our customer service team by email at info@weberscientific.com or by phone at 800-328-8378.

How to Order Air Sampler Recalibration

1. Order Air Sampler Recalibration item **4140-93** from Weber Scientific. Please see **weberscientific.com** for current pricing information and to order.
2. Send the following items to the address below:
 - A completed order form (below)
 - Your carefully packaged Air Sampler

IEH Engineering
 13748 32nd Ave NE, Unit B
 Seattle, WA 98125
 Attn: AS Calibration

3. Send a copy of this form to **info@weberscientific.com** Form must be received by email before recalibration can be performed.

AIR SAMPLER RECALIBRATION ORDER FORM

Customer Information:

Name : _____ Weber Account # : _____ Date : _____

Billing Information:	
Company	
Address 1	
Address 2	
City, State, Zip	
Contact	
Phone	
Email	

Shipping Information (if different than billing):	
Company	
Address 1	
Address 2	
City, State, Zip	
Contact	
Phone	
Email	

Air Sampler Information:

Serial Number(s):

Return Shipment:

- Ground
 2 Day
 UPS Collect Acct # _____
 3 Day
 Next Day
 Fedex Collect Acct # _____